

**76 titled references relevant to
MRI instrumentation up to the Spring of 1984**

Collected by Stanislav Sykora
while preparing the Italian book

Risonanza Magnetica Nucleare in Medicina

Ruffato C., Bonera G., Sykora S., Buttazzoni L., Paolucci G.
Piccin Nuova Libreria, Padova 1984,
([http://www.ebyte.it/stan/SS Papers.html#31](http://www.ebyte.it/stan/SS_Papers.html#31))

Parte IV:

Principi fisici di formazione di immagine tramite RMN.

URL of this file:

[www.ebyte.it/library/refs/1984 Sykora MRI-Instrumentation Refs.pdf](http://www.ebyte.it/library/refs/1984_Sykora_MRI-Instrumentation_Refs.pdf)

Other works by Stan Sykora can be found at
www.ebyte.it

1. Abragam A. The principles of nuclear magnetism. Clarendon press, Oxford, 1961
2. Adduci D.J./Hornung P.A./Torgeson D.R. Extremely fast recovery time receiver for pulsed NMR. *Rev. Sci. Instrum.* 47: 1503, 1976
3. Alderman D.W./Grant D.M. Efficient decoupler coil design which reduces heating in conductive samples in superconducting spectrometers. *J. Mag. Res.* 36: 447, 1979
4. Anderson W.A. Electrical current shims for correcting magnetic fields. *Rev. Sci. Instrum.* 32: 241, 1961
5. Assink R.A. A quadrupole coil for nuclear magnetic resonance spin-echo diffusion measurements as a function of pressure. *J. Magn. Res.* 22: 165, 1976
6. Baines T./Mansfield P. Improved picture display for NMR imaging. *J. Phys. E*: 9, 809, 1976
7. Bangert V./Mansfield P. Magnetic field gradient coils for NMR imaging. *J. Phys. E*: 15, 1982
8. Battocletti J.H./Halbach R.E./Salles-Cunha S.X./Sances A.JR. The NMR blood flow meter – theory and history. *Med. Phys.* 8: 435, 1981
9. Battocletti J.H./Halbach R.E./Sances A.Jr./Larson S.J./Bowman R.L./Kudravcev V. Flat crossed coil detector for blood flow measurements using nuclear magnetic resonance. *Med. Biol. Eng. Comput.* 17: 183, 1979
10. Becker J.J. Permanent magnets. *Sci. Am.* 223: 92, 1970
11. Bergmann W.H. A new approach to NMR flow imaging and analysis. In « NMR imaging ». Ed. Partain C.L./et al., W.B. Saunfers C., Philadelphia, 1983
12. Blicharski J.S./Sobol W.T. A new type of magnetic field gradient coil for NMR measurements. *J. Magn. Res.* 46: 1, 1982
13. Bottomley P.A. A versatile magnetic-field gradient control system for NMR imaging. *J. Phys. E* 14: 1081, 1981
14. Bottomley P.A. Digital gradient magnetic-field reorientation in 3-dimensional NMR zeumatogra phy. *J. Phys. E* 14: 1052, 1981
15. Bottomley P.A./Hinshaw W.S./Holland G.N. Computer driven photo scanner for medical imaging. *Phys. Med. Biol.* 23: 309, 1978
16. Bowman R.L./Kudravcev V. Bloo flowmeter utilizing nuclear magnetic resonace. *Ire Trans. Med. Electron* 6: 267, 1959
17. Casselman T.N./Fryer R.B./Simon F.N. Uniformity condition for a rectangular field coil with a split winding. *Rev. Sci. Instrum.* 49: 80, 1978
18. Colavita P.A./Bustos V.A. Optimum spacing for a pair of thick circular coils of square section. *Rev. Sci. Instrum.* 49: 1006, 1978
19. Colligiani A./Ambrosetti R./De Lisi P. A microfonicity-free Robinson-type NQR spectrometer using dual-gate mosfets. *J. Magn. Res.* 20: 341, 1975
20. Crooks L.E./Arakawa M./Hoenninger J./Watts J./McRee R./Kaufman L./Davis P.L./Margulis A.R./De Groot J. Nuclear magnetic resonance whole-body imager operating at 3.5 kGauss. *Radiology* 143: 169, 1982
21. Crownfield F.R.Jr. Optimum spacing of coil pairs. *Rev. Sci. Instrum.* 35: 240, 1964
22. Edelstein W.A./Bottomley P.A./Pfeifer L.M. A signal-to-noise calibration procedure for NMR imaging systems. *Med. Phys.* 11 (2): 180, 1984
23. Fitzsimmons J.R. Gradient control system for nuclear magnetic resonance imaging. *Rev. Sci. Instrum.* 53 (9): 1338, 1982
24. Gadian D.G./Robinson F.N. Radiofrequency losses in NMR experiments on electrically conducting samples. *J. Magn. Res.* 34: 449, 1979
25. Garrett M.W. Axially symmetric systems for generating and magnetic measuring fields. *J. Appl. Phys.* 22: 1091, 1951
26. Geballe T.H./Hulm J.K. Superconductors in electric power technology. *Sci. Am.* 243: 138, 1980
27. Gilbert E./Hanley P./Hayter J.C./White J.W. An asymmetric split-pair superconducting magnet for nuclear polarization experiments. *J. Phys. E* 6: 714, 1973
28. Ginsberg D.M./Melchner M.J. Optimum geometry of saddle shaped coils for generating a uniform magnetic field. *Rev. Sci. Instrum.* 41: 122, 1970
29. Golay M.J.E. U.S. patent No. 3569523 and 3622869, 1971
30. Golay M.J.E. Field homogenizing coils for nuclear spin resonance instrumentation. *Rev. Sci. Instrum.* 29: 313, 1958
31. Gordon R.E./Timms W.E. An improved tune and match circuit for be shimming in intact biological samples. *J. Mag. Res.* 46: 322, 1982
32. Grisenti R./Zecca A. Design data for a triple square coil system. *Rev. Sci. Instrum.* 52: 1097, 1981
33. Gubser D.U./Wolf S.A./Cox J.E. Shielding of longitudinal magnetic fields with thin, closely spaced, concentric cylindres of high permeability. *Rev. Sci. Instrum.* 50: 751, 1979
34. Hilal M.A./Eyssa Y.M. Generating uniform fields in the interior region of different shaped

- circular dipoles. *Ieee Trans. On Mag. Mag-18*: 688, 1982
35. Holland G.N./Bottomley P.A. Color display technique for NMR imaging. *J. Phys E* 10: 714, 1977
 36. Holland G.N./Heysmond E. Solid-state high-power rf-amplifier for pulsed NMR. *J. Phys. E* 12: 480, 1979
 37. Hoult D.I. Fast recovery, high sensitivity NMR probe and preamplifier for low frequencies. *Rev. Sci. Instrum.* 50: 193, 1979
 38. Hoult D.I./Chen C.N./Eden H./Eden M. Elimination of baseline artifacts in spectra and their integrals. *J. Magn. Res.* 51: 110, 1983
 39. Hoult D.I./Lauterbur P.C. Sensivity of the zeugmatographic experiment involving human samples. *J. Magn. Res.* 34: 425, 1979
 40. Hoult D.I./Richards R.E. The signal-to-noise ratio of the nuclear magnetic resonance experiment. *J. Magn. Res.* 24: 71, 1976
 41. Hoult D.I./Richards R.E. A VHF fet preamplifier with 0.3 dB noise figure electron. *Lett.* 11: 596, 1975
 42. Hoult D.I./Richards R.E. Critical factors in the design of sensitive high resolution nuclear magnetic resonance spectrometers. *Proc. Roy. Soc. London A* 344: 311, 1975
 43. Hulm J.K./Kunzler J.E./Matthias B.T. The road to superconducting materials. *Phys. Today* Jan.: 34, 1981
 44. Hulm J.K./Matthias B.T. High field, high current superconductors. *Science* 288: 881, 1980
 45. Hutchison J.M.S./Sutherland R.J./et al. Coil winding for quadrupolar fields. *Uk patent* 7934864
 46. Jackson J.A. Whole body NMR spectrometer. *Rev. Sci. Instrum.* 39: 510, 1968
 47. Johnson G./Hutchinson J.M.S./Eastwood L.M. Instrumentation for NMR spin-warp imaging. *J. Phys. E* 15: 74, 1982
 48. Kan S./Fan M./Courtieu J. A simple-coil triple resonance probe for NMR experiments. *Rev. Sci. Instrum.* 51 (7): 887, 1980
 49. Katyl R.H. FFT calculation of magnetic fields in arit coils. *Ieee Trans. On Mag. Mag-16*: 545, 1980
 50. Katyl R.H. Air-coil design using numerical optimization methods. *Ieee Trans. On Mag. Mag-16*: 1098, 1980
 51. Kaufman L./Shose D.W. Generalized methodology for the comparison of diagnostic imaging instrumentation. *Afips Nat. Comput. Conf. Expo Cory Proc.* 49: 445, 1980
 52. Kolm H.H./Freeman A.J. Intense magnetic fields. *Sci. Am.* 212: 66, 1965
 53. Kroon D.J. Laboratory magnets. *N.V. Phillips Gloeilampenfabrieken*, Eindhoven, Netherlands, 1968
 54. Kunzler J.E./Tanenbaum M. Superconducting magnets. *Sci. Am.* 206: 60, 1952
 55. Lai C.M./Lauterbur P.C. A gradient control device for complete three-dimensional nuclear magnetic resonance zeugmatographic imaging. *J. Phys. E* 13: 747, 1980
 56. Lai C.M./Shook J.W./Lauterbur P.C. Micro-processot-controlled reorientation of magnetic-field gradients. *Chem. Biomed. Environ. Instrum.* 9: 1, 1979
 57. Lindon J.C./Ferrige A.G. Digitization and data processing in fourier transform NMR. *Progress in NMR spectroscopy* 14: 27, 1980 Ed. Emsley J.W./Feeney J./Sutcliffe L.H., Pergamon Press, Oxford.
 58. Lowe I.J./Tarr C.E. A fast recovery probe and receiver for pulsed nuclear magnetic resonance spectroscopy. *J. Phys. E* 1: 320, 1968
 59. Mallard J.R./Hutchison J.M.S./Edelstein W.A./Ling C.R./Foster M.A./Johnson G. in vivo NME imaging in medicine: the Aberdeen approach, both physical and biological. *Phil. Trans. R. Soc. London* B289: 519, 1980
 60. Mansfield P./Morris P.G. NMR imaging in biomedicine in advances in magnetic resonance. Ed. Waugh J.S., Academy Press, New York, 1982
 61. Merritt R./Purcell C./Stroink G. Uniform magnetic field produced by three, and five square coils. *Rev. Sci. Instrum.* 54: 879, 1983
 62. Montgomery D.B. Solenoid magnet design: the magnetic and mechanical aspects of resistive and superconducting systems. *Wiley-Interscience*, New York, 1969
 63. Murgatroyd P.N. Optimum designs of circular coil system for generating nonuniform axial magnetic fields. *Rev. Sci. Instru.* 50: 668, 1979
 64. Oedberg G./Oedberg L. On the use of a quadrupole coil for NMR spin-echo diffusion studies. *J. Mag. Res.* 16: 342, 1974
 65. Parker R.S./Zupancic I./Pirs J. Coil system to produce orthogonal, linear magnetic field gradients. *J. Phys. E* 6: 899, 1973
 66. Parks R.D. (Ed.) *Superconductivity*. Marcel Dekker Inc., New York, 1969
 67. Pasotti G./Sacchetti T. Tecnologia dei materiali superconduttori per elevati campi magnetici. *CNEN-LNF* 68 / 7: 1968, Frascati
 68. Redpath T.W./Hutchison J.M.S. Design of a radiofrequency coil suitable for NMR imaging of heads. *Phys. Med. Biol.* 27: 443, 1982
 69. Rudd M.E./Craig J.R. Optimum spacing of square and circular coil pairs. *Rev. Sci. Instrum.* 39: 1372, 1968
 70. Shosa D.W./Kaufman L. Methods for evaluation of diagnostic imaging instrumentation. *Phys. Med. Biol.* 26: 101, 1981
 71. Simon H.E. A whole body NMR imaging system with full three-dimensional capabilities.

- Proc. Soc. Photo-opt. Instru. Eng.* 273: 41, 1981
72. Sykora S./Vogt J. Experimental noise in data acquisition and evaluation. II. Signal to noise ratio measurements in NMR. *Bruker Report* 2/79: 17, 1979
 73. Sykora S./Vogt J. Experimental noise in data acquisition and evaluation. I. Noise amplitude distrib. function on HR-FT-NMR instrum. *Bruker Report* 2/79: 15, 1979
 74. Tanner J.E. Pulsed field gradients for NMR spin-echo diffusion measurements. *Rev. Sci. Instr.* 36: 1086, 1965
 75. Webster D.S./Marsden K.H. Improved apparatus for the NMR measurement of selfdiffusion coefficients using pulsed field gradients. *Rev. Sci. Instrum.* 45: 1232, 1974
 76. Zupancic I./Pris J. Coils producing a magnetic field gradient for diffusion measurements with NMR. *J. Phys. E* 9: 79, 1976